



Calibration Cert# 2166.01



Certificate of Calibration

Everett Service Center

Certificate Number:	EVL743342	Calibration Date:	01-Oct-2021
Data Type:	As-Left	Certificate Date:	01-Oct-2021
Result Summary:	Measurement Results < Limits	Temperature:	22.5 °C
Manufacturer:	Fluke	Humidity:	44.1 %
Model:	525B		
Serial Number:	2041092		
Description:	Temperature / Pressure Calibrator		

Procedure:	Fluke 525A_B (1Year_90 Day) Calibration Verification	Revision:	1.0
Customer:	ILLIANA INSTRUMENTATION INCORPORATED		
City:	SCHERERVILLE	Country:	US
State:	IN		
Purchase Order:	4236	RMA:	32279259
Asset ID:	1546		

This calibration is traceable to the International System of Units (SI), through National Metrology Institutes (NIST, PTB, NRC, NPL, etc.), radiometric techniques, or natural physical constants. This certificate applies only to the item identified and shall not be reproduced other than in full, without the specific written approval by Fluke Corporation. Calibration certificates without signature are not valid. The calibration has been completed in accordance with Fluke Electronics Corporation Quality System Document 111.0 Revision 124 and/or Fluke 17025 Quality Manual QSD 111.41 Revision 007.

The Data Type found in this certificate must be interpreted as:

- As - Found Calibration data collected before the unit is adjusted and / or repaired.
- As - Left Calibration data collected after the unit has been adjusted and / or repaired.
- Found-Left Calibration data collected without any adjustment and / or repair performed.

This calibration conforms to the requirements of ISO/IEC 17025:2017 and ANSI/NCCL Z540-1-1994 (R2002).


In the attached measurement results, deviation may be expressed with units, Measured Value (MV) - Nominal Value (NV) or as a proportion of the nominal value ((MV-NV)/NV), expressed without units with a scalar multiplier such as % (0.01), or as a ratio of the units (mA/A, μV/V, etc.) Descriptions such as μA/A, μV/V, and others, where used to annotate results or column headings are the preferred replacements for what was historically labeled as "ppm" or parts-per-million and described the results in that column, unless otherwise noted by units symbols.

Where applicable, the expanded uncertainty of measurement at the time of test is given in the following pages. They are calculated in accordance with the method described in the ISO Guide to the Expression of Uncertainty in Measurement (GUM). The reported expanded uncertainty of measurement is stated as the standard uncertainty of measurement multiplied by the coverage factor k, such that the confidence level approximates 95%.

This calibration certificate may contain data that is not covered by the A2LA Scope of Accreditation. Unaccredited material, where applicable is indicated by an asterisk (*), or confined to clearly marked sections. Functional (Pass / Fail) tests are not accredited.

No statement of compliance with specifications is made or implied on this certificate. However, measurement results are reviewed, where applicable, to establish where any measurement result exceeded the manufacturer's specifications.

Measurement results greater than limits of error are indicated by '!'.



William Callaghan
 William Callaghan
 Calibration Technician

Certificate Number: EVL743342**Date of Calibration:** 01-Oct-2021**Standards Used**

Asset	Description	Cal-Date	Cal-Due
13861	ESI SR1030 Resistance Transfer Standard	10-Jun-2021	10-Oct-2021
13862	ESI SR1030 Resistance Transfer Standard	11-Feb-2021	11-Feb-2022
10776	Fluke 5520A Calibrator	15-Mar-2021	15-Dec-2021
15765	Fluke 5640 Thermistor Probe	03-Sep-2021	03-Sep-2022
7668	Fluke 742A-1 Resistance Standard	09-Dec-2020	09-Dec-2021
J1713	FLUKE 8508A Reference Multimeter	08-Dec-2020	08-Dec-2021
J1287	Hart Scientific 1504 Thermometer Readout	06-Apr-2021	06-Apr-2022
14367	Omega TJ36-ICIN-18U-12-SB-SMPW-M "J" Type Rugged Transition Joint Probe	27-Apr-2021	27-Apr-2022

Certificate Number: EVL743342

Date of Calibration: 1-Oct-2021

Calibration Data

Parameter	Nominal Value	Measurement Result	Limits of Error		Expanded Uncertainty
			Lower Limit	Upper Limit	
INSTRUMENT IDENTIFICATION					
Firmware Version: V1.31					
1 Year Specifications shown unless otherwise noted					
DC VOLTAGE OUTPUT:					
100mV RANGE					
0.000000 V	0.000000	0.0000014	-0.0000030	0.0000030	8.0e-007 V
0.025000 V	0.025000	0.0250011	0.0249963	0.0250038	8.4e-007 V
0.075000 V	0.075000	0.0750006	0.0749947	0.0750052	1.3e-006 V
0.100000 V	0.100000	0.1000010	0.0999940	0.1000060	1.2e-006 V
1 VOLT RANGE					
0.00000 V	0.00000	0.000002	-0.000010	0.000010	5.8e-006 V
0.25000 V	0.25000	0.250005	0.249983	0.250018	6.2e-006 V
0.75000 V	0.75000	0.750009	0.749968	0.750032	7.3e-006 V
1.00000 V	1.00000	1.000014	0.999960	1.000040	7.7e-006 V
10 VOLT RANGE					
0.0000 V	0.0000	0.00001	-0.00010	0.00010	5.8e-005 V
2.5000 V	2.5000	2.50002	2.49983	2.50018	6.2e-005 V
7.5000 V	7.5000	7.50002	7.49967	7.50033	7.3e-005 V
10.0000 V	10.0000	10.00005	9.99960	10.00040	7.7e-005 V
100 VOLT RANGE					
0.000 V	0.000	0.0002	-0.0010	0.0010	5.8e-004 V
25.000 V	25.000	25.0000	24.9982	25.0018	6.2e-004 V
75.000 V	75.000	74.9994	74.9968	75.0032	7.4e-004 V
100.000 V	100.000	99.9994	99.9960	100.0040	8.2e-004 V
DC CURRENT OUTPUT					
100mA Range					
0.00000 mA	0.00000	0.000001	-0.001000	0.001000	5.9e-009 A
25.00000 mA	25.00000	24.999695	24.997750	25.002250	2.6e-007 A
75.00000 mA	75.00000	74.999219	74.995250	75.004750	3.0e-007 A
100.00000 mA	100.00000	99.99918	99.99400	100.00600	4.3e-007 A
THERMOCOUPLE INPUT					
-5.000 mV	-5.0000	-5.001	-5.003	-4.997	1.0e-006 V
15.000 mV	15.0000	15.000	14.997	15.003	1.2e-006 V
30.000 mV	30.0000	30.000	29.996	30.004	1.6e-006 V
50.000 mV	50.0000	50.000	49.996	50.005	1.7e-006 V

Calibration Data

Parameter	Nominal Value	Measurement Result	Limits of Error		Expanded Uncertainty
			Lower Limit	Upper Limit	
70.000 mV	70.0000	70.000	69.995	70.005	1.9e-006 V
THERMOCOUPLE OUTPUT					
-5.000 mV	-5.000	-4.9996	-5.0031	-4.9969	6.5e-007 V
15.000 mV	15.000	14.9996	14.9965	15.0035	6.1e-007 V
30.000 mV	30.000	30.0003	29.9961	30.0039	6.3e-007 V
50.000 mV	50.000	50.0003	49.9955	50.0045	6.7e-007 V
70.000 mV	70.000	70.0000	69.9949	70.0051	7.2e-007 V
RESISTANCE OUTPUT					
400 Ohm Range					
5.000 Ohm	5.000	4.9996	4.9850	5.0150	6.4e-004 Ω
100.000 Ohm	100.000	100.0004	99.9850	100.0150	1.1e-003 Ω
200.000 Ohm	200.000	200.0030	199.9850	200.0150	2.4e-003 Ω
300.000 Ohm	300.000	300.0049	299.9850	300.0150	3.0e-003 Ω
400.000 Ohm	400.000	400.0077	399.9850	400.0150	3.7e-003 Ω
4000 Ohm Range					
5.0 Ohm	5.0	5.01	4.70	5.30	5.8e-002 Ω
1000.00 Ohm	1000.00	999.987	999.700	1000.300	9.9e-003 Ω
2000.00 Ohm	2000.00	2000.000	1999.700	2000.300	2.6e-002 Ω
3000.00 Ohm	3000.00	3000.002	2999.700	3000.300	3.0e-002 Ω
4000.00 Ohm	4000.00	4000.018	3999.700	4000.300	3.7e-002 Ω
RESISTANCE MEASURE					
400 Ohm Range					
0.00 Ohm	0.00000	0.0003	-0.0040	0.0040	1.2e-003 Ω
100.00 Ohm	100.0045	100.004	99.998	100.010	1.2e-003 Ω
200.00 Ohm	200.0059	200.004	199.998	200.014	1.2e-003 Ω
300.00 Ohm	300.0082	300.006	299.998	300.018	1.2e-003 Ω
400.00 Ohm	400.0109	400.007	399.999	400.023	1.2e-003 Ω
4000 Ohm Range					
0.00 Ohm	0.0000	0.000	-0.030	0.030	9.7e-004 Ω
1000.00 Ohm	999.586	999.59	999.53	999.65	1.1e-002 Ω
2000.00 Ohm	1999.219	1999.17	1999.14	1999.30	1.1e-002 Ω
3000.00 Ohm	2998.854	2998.81	2998.75	2998.95	5.9e-003 Ω
4000.00 Ohm	3998.484	3998.43	3998.36	3998.60	5.9e-003 Ω
COLD JUNCTION COMPENSATION					
22.7545 °C	22.755	22.78	22.59	22.91	6.0e-002 °C

Certificate Number: EVL743342

Date of Calibration: 1-Oct-2021

Calibration Data

Parameter	Nominal Value	Measurement Result	Limits of Error		Expanded Uncertainty
			Lower Limit	Upper Limit	
PRESSURE MODULE IDENTIFICATION					
Pressure module properly identified		Pass			