

# Certificate of Calibration



Illiana Instrumentation Service LLC  
1831 Govert Drive  
Scherverville, IN 46375



Cert #: 11464 Rev. 1 Item # 1718 Certificate Date: 7/15/2023

## CUSTOMER

ILLIANA INSTRUMENTATION SERVICE LLC  
1831 GOVERT DRIVE  
SCHERERVILLE, IN 46375

Purchase Order:

## EQUIPMENT INFORMATION

Item # 1718 Calibration Result: PASS  
Manufacturer HANNA INSTRUMENTS INC.  
Model Number HI9564  
Serial Number A04090149111  
Equip. Location LAB/MOBILE  
Description DIGITAL THERMOHYGROMETER

## CALIBRATION INFORMATION

Date received: 7/15/2023 Cal Date: 7/15/2023 Cal. Interval 12 MONTHS Cal. Due Date: 7/15/2024  
Technician: PAUL GROLLA Temperature: 70 F Humidity 50 Cal location LAB  
Tag

Department:

Notes:

Problems: None

Adjustments None

Equipment Notes

Sensitivity OKAY

Initial modification offset 0

Final modification offset 0

Note: If the modification offset is not zero, readings reflect zero

Limitations NONE

Accuracy +/- 1 DEG C./5% RH

## Quality and traceability statements

This certifies that the above equipment was calibrated using appropriate Illiana Instrumentation technical procedures. At planned intervals, Illiana Instrumentation standards are calibrated by comparison to or measurement against standards which are traceable to the SI units through the NIST or other recognized national measurement institutes or international standard bodies. The results in this report relate only to the item(s) calibrated. If so indicated above, detailed calibration results are attached to this certificate. These results are part of this certificate and this certificate shall not be reproduced except in full, without the written approval of Illiana Instrumentation. Any number of factors not under the control of the calibration laboratory may cause the calibration of the above item(s) to drift before the recommended recalibration date. Supporting documentation relative to traceability and technical procedures used is on file and is available for examination upon request and approval of our quality assurance manager. The above uncertainties represent an expanded uncertainty expressed at approximately 95% confidence level using a coverage factor of k=2. The certificate date on this report constitutes the issue date. Pass/Fail criteria does not take into account measurement uncertainty.

Calibrated By:

*Paul Grolla*

Approved By:

Laura Grolla

Customer Approval \_\_\_\_\_



Illiana Instrumentation Service LLC  
 1831 Govert Drive  
 Schererville, IN 46375



Cert #: 11464 Rev. 1 Item # 1718 Certificate Date: 7/15/2023

Calibration Data

Values found OOT and adjusted are indicated by "A" and items left OOT are indicated by "F"

Group	Description	Nominal	Lower Tol.	Upper Tol.	As Found	As Left	Units	Uncert.	OOT
	% rH	33.0	28.0	38.0	38	38	%RH	2.5	
	% Rh	74.2	63.1	73.1	72	72	%RH	2.5	
	Deg C.	30.0	29.0	31.0	30	30	°C	0.8	
	Deg C.	23.0	22.0	24.0	23	23	°C	0.8	

Standards Used

I.D.	Model	Description	Manufacturer	Serial No.
1653	HYGROPALM HP22-A	THERMOHYGROMETER	ROTRONIC	61626838

Procedures Used

Name	Description
WI-520	HUMIDITY PROBE CALIBRATION