



CERTIFICATE OF ANALYSIS
Complies with ISO 17034, ISO Guide 31,
ISO Guide 35, and ISO 9001
TRACEABLE® CERTIFIED REFERENCE MATERIAL



This certificate indicates traceability to standards provided by (NIST) National Institute of Standards and Technology and/or a National Standards Laboratory.

1728

Certificate No. : 4889-12014827
Description : Buffer 10.012 pH Blue
Catalog Number : 15-078-287,15568703 **Lot :** CC715120
Certificate Date : 03 Mar 2021 **Expiration Date :** 03 Mar 2023
Certified Value : 10.014 pH **U:** ±0.011 pH (k=2) at 25°C

Certification measurements are performed under ISO 17034, A2LA accreditation no. 1750.02. They are traceable to recognized national and international standards via an unbroken chain of comparisons. pH is defined as the negative logarithm of the hydrogen ion activity.

MEASUREMENT: Minimum twelve (12) 100 ml samples were measured from this lot. The pH of each sample was determined using a pH meter and electrode.

UNCERTAINTY: The certified value is given as the average of the measured samples. The reported expanded uncertainty (U) is determined from the measurement variation from sample to sample, change due to shelf life, and uncertainty of the measurement process. The value of uncertainty is multiplied by k=2, which for a normal distribution corresponds to a coverage probability of approximately 95%. Uncertainty is calculated in accordance with the ISO "Guide to the Expression of Uncertainty in Measurement".

METHOD: The certified reference material is produced and analyzed by Control Company. The certified reference material is comprised of deionized water, 0.025 mol/kg H₂O sodium hydrogen carbonate, and 0.025 mol/kg H₂O sodium carbonate.

Marisa Elms

Marisa Elms, Technical Manager

Nicol Rodriguez

Nicol Rodriguez, Quality Manager

Traceability: Standards and Equipment Used

Description	Serial Number	Due Date	Traceable Reference
pH/Ion Meter	658R067-N023		
pH Reference Material 10.000	E1404120	25 Mar 2023	041-20
Digital Thermometer	102008443	22 Feb 2022	4000-11972359
pH Electrode	15293-F18		

Laboratory Environment Conditions: 40.00%RH 24.80°C 1018mBar

CONTROL COMPANY 12554 Galveston RD Suite B230 Webster TX USA 77598
Phone 281 482-1714 Fax 281 482-9448 sales@control3.com www.traceable.com

Control Company is an ISO 17034:2016 Certified Reference Material (CRM) Producer Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.02). This certificate fulfills the requirements of ISO Guide 31:2015 (Reference Materials – Contents of Certificates and Labels), ISO 17034:2016 "Quality System Guidelines for the Production of Reference Materials", and ISO Guide 35:2017 "Certification of Reference Materials – General and Statistical Principles". Control Company is an ISO/IEC 17025:2017 Calibration Laboratory Accredited by American Association for Laboratory Accreditation (A2LA Certificate No. 1750.01). Control Company is ISO 9001:2015 certified by DNV GL (Certificate No. CERT-01805-2006-AQ-HOU-ANAB). Traceable® is a registered trademark of Control 3 Inc.