

Item 1682



AMETEK Brookfield
INSTRUMENTATION & SPECIALTY CONTROLS DIVISION
11 Commerce Blvd. | Middleboro, MA 02346
P: 508.946.6200 | F: 508.946.6262
www.brookfieldengineering.com

CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

Viscosity Standard: 5000		Lot Number: 091819	
Certification/Issue Date: 091819		Expiry Date: March 30, 2021	
Viscosity		Temperature	
(centipoise)		(°C)	(°F)
4820		25.0	77.00
Model	TEST INSTRUMENT	Serial Number	
LVDV-II+		6524090	
Model	TEMPERATURE SENSOR	Serial Number	
1502A		B0C663	

CALIBRATION OF THE ABOVE MENTIONED AMETEK BROOKFIELD VISCOSITY STANDARD IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.
THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH AN AMETEK BROOKFIELD VISCOMETER THAT HAS BEEN CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

APPROVAL SIGNATURE

David J. Moonay, Ph.D -Quality Engineer - Rheologist