

# ITEM 1683



**AMETEK Brookfield**  
**INSTRUMENTATION & SPECIALTY CONTROLS DIVISION**  
 11 Commerce Blvd. | Middleboro, MA 02346  
 P: 508.946.6200 | F: 508.946.6262  
 www.brookfieldengineering.com

## CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

<i>Viscosity Standard: 1000</i>		<i>Lot Number: 091219</i>	
<i>Certification/Issue Date: 09/12/2019</i>		<i>Expiry Date: April 02, 2021</i>	
<b>Viscosity</b>		<b>Temperature</b>	
(centipoise)		(°C)	(°F)
945		25.0	77.00
<b>Model</b>	<b>TEST INSTRUMENT</b>		<b>Serial Number</b>
LVDV-II+			6529213
<b>Model</b>	<b>TEMPERATURE SENSOR</b>		<b>Serial Number</b>
1529			A56933

CALIBRATION OF THE ABOVE MENTIONED AMETEK BROOKFIELD VISCOSITY STANDARD  
 IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE.  
 THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH AN AMETEK BROOKFIELD VISCOMETER THAT HAS BEEN  
 CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

APPROVAL SIGNATURE

David J. Moonay, Ph.D -Quality Engineer - Rheologist