

ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Illiana Instrumentation
1831 Govert Drive
Schererville, IN 46375-

Date received: 10/25/18
Calibration date: 10/25/18
Recalibration due: 10/31/19

Purchase order:
Location of calibration: Lab

Item number: 1278	Item location: Lab
Tag number:	Temp: 72.0° F. Humidity: 62 Sensitivity: OK
Manufacturer: Ametek	Condition as found: In tolerance
Model num: C-140	Condition as left: In tolerance
Description: 1278 Dry Block	Intentional offset as found: 0.0000
Serial number: 015123-00130	Intentional offset as left: 0.0000
Accuracy: +/- .5 Deg. C.	Limitations: None
Item notes:	

This certifies that the above equipment was calibrated using appropriate Illiana Instrumentation technical procedures. At planned intervals, Illiana Instrumentation standards are calibrated by comparison to or measurement against standards which are traceable to the SI units through the NIST or other recognized national measurement institutes or international standard bodies. The results in this report relate only to the item(s) calibrated. If so indicated above, detailed calibration results are attached to this certificate. These results are part of this certificate and this certificate shall not be reproduced except in full, without the written approval of Illiana Instrumentation. Any number of factors not under the control of the calibration laboratory may cause the calibration of the above item(s) to drift before the recommended recalibration date. Supporting documentation relative to traceability and technical procedures used is on file and is available for review upon request and approval of our quality assurance manager. The above uncertainties represent an expanded uncertainty expressed at approximately 95% confidence level using a coverage factor of k=2.

Tech: Paul D Grolla Date: 10/25/18 Approved by: Laura Grolla Date: 10/25/18
Paul D Grolla - ISA Certified Controls Systems Tech. Laura Grolla - Deputy Quality Manager

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
20.0000	20.0200	20.0200	20.5000	19.5000	Deg. C.		Yes

Standard(s): 1225 Hart PRT, 1429 Hart Scanner

Adjustments:

Problems:

Uncertainty: 0.2100 Deg. C.

Procedure: WI-516

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
60.0000	60.0370	60.0370	60.5000	59.5000	Deg. C.		Yes

Standard(s): 1225 Hart PRT, 1429 Hart Scanner

Adjustments:

Problems:

Uncertainty: 0.2100 Deg. C.

Procedure: WI-516

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
-15.0000	-14.9160	-14.9160	-14.5000	-15.5000	Deg. C.		Yes

Standard(s): 1225 Hart PRT, 1429 Hart Scanner

Adjustments:

Problems:

Uncertainty: 0.2100 Deg. C.

Procedure: WI-516

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
100.0000	100.0360	100.0360	100.5000	99.5000	Deg. C.		Yes

Standard(s): 1225 Hart PRT, 1429 Hart Scanner

Adjustments:

Problems:

Uncertainty: 0.2100 Deg. C.

Procedure: WI-516



ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 (219)227-8788 illianaservice.com

CERTIFICATE OF CALIBRATION

Cal Point	Init Read	Final Read	Upper Tol	Lower Tol	Eng Units	Cal point description	Pass?
140.0000	140.0440	140.0440	140.5000	139.5000	Deg. C.		Yes

Standard(s): 1225 Hart PRT, 1429 Hart Scanner

Adjustments:

Problems:

Uncertainty: 0.2100 Deg. C.

Procedure: WI-516

Customer approval: _____ Date: _____

Total pages for Certificate 999999-01278-5430: 2