



ILLIANA INSTRUMENTATION
 1831 Govert Drive Schererville, IN 46375 Phone (219)227-8788 Fax (219)515-6161
CERTIFICATE OF CALIBRATION

<p style="text-align: center;">CUSTOMER:</p> <p>Illiana Instrumentation Service 1831 Govert Drive Schererville, IN 46375</p>	<p style="text-align: center;">MISCELLANEOUS DETAILS:</p> <p>Date Received 10/19/18 Certification Date: 10/19/18 Recalibration Date: 1/15/19 Cal. Number: 1468-101918 P.O. Number: Location of Calibration: Lab Detailed Results Attached: YES Procedure Used: Fluke Procedure</p>
---	--

EQUIPMENT CALIBRATED	
MANUFACTURER:	Fluke
MODEL:	725
SERIAL NUMBER:	1100083
ITEM NUMBER:	1468
DESCRIPTION:	Calibrator
CONDITION AS FOUND:	In tolerance

STANDARDS USED	
Item 1546, Fluke 525B; Item 1205 HP34401A	

TEST CONDITIONS	
TEMPERATURE	75 Deg F.
HUMIDITY	37% rH

CERTIFIED BY: Paul Drath TITLE: ISA CCST III DATE: 10/19/18
 APPROVED BY: Laura Giolla TITLE: Op Mgr. DATE: 10/19/18

This certifies that the above equipment was calibrated using appropriate Illiana Instrumentation technical procedures. At planned intervals, Illiana Instrumentation standards are calibrated by comparison to or measurement against standards which are traceable to the SI units through the NIST or other recognized national measurement institutes or international standard bodies. The results in this report relate only to the item(s) calibrated. If so indicated above, detailed calibration results are attached to this certificate. These results are part of this certificate and this certificate shall not be reproduced except in full, without the written approval of Illiana Instrumentation. Any number of factors not under the control of the calibration laboratory may cause the calibration of the above item(s) to drift before the recommended recalibration date. Supporting documentation relative to traceability and technical procedures used is on file and is available for examination upon request and approval of our quality assurance manager. The above uncertainties represent an expanded uncertainty expressed at approximately 95% confidence level using a coverage factor of k=2.

Item	1468 Fluke 725								
Accuracy:	Varies By Range								
Date:	10/19/2018								
Intentional Offset as Found	None								
Intentional Offset as Left	None								
Limitations	None								
Input range	Eng. Units	Cal Pt	Upper	lower	Initial	Final	Sensitivity	STD	Unc.
							Check		
Meas VDC Upper Display	VDC	0	0.002	-0.002	0	0	OK	1546	0.000033
		15	15.005	14.995	15	15	OK	1546	0.005
		30	30.008	29.992	29.999	29.999	OK	1546	0.005
Meas mVDC on lower display	mDVC	0	0.02	-0.02	0	0	OK	1546	0.0033
		45	45.03	44.97	44.99	44.99	OK	1546	0.0033
		90	90.04	89.96	89.99	89.99	OK	1546	0.0033
Lower display voltage measure	VDC	0	0.002	-0.002	0	0	OK	1546	0.000033
		10	10.004	9.996	9.998	9.998	OK	1546	0.005
		20	20.006	19.994	19.996	19.996	OK	1546	0.005
Meas mADC 30 mADC upper display	mADC	4	4.003	3.997	3.999	3.999	OK	1546	0.0016
		12	12.005	11.995	11.999	11.999	OK	1546	0.0016
		24	24.007	23.993	23.998	23.998	OK	1546	0.0057
Meas mADC 30 mADC lower display	mADC	4	4.003	3.997	3.999	3.999	OK	1546	0.0016
		12	12.005	11.995	11.998	11.998	OK	1546	0.0016
		24	24.007	23.993	23.996	23.996	OK	1546	0.0057
Frequency source lower display	kHz	10	10.025	9.975	9.9998	9.9998	OK	1205	0.0033
Lower display 4 wire resistance measure	Ohms	15	15.1	14.9	14.97	14.97	OK	1546	0.025
		350	350.1	349.9	349.99	349.99	OK	1546	0.025
		500	500.5	499.5	499.9	499.9	OK		0.049
Lower display 3 wire RTD measure	Ohms	350	350.2	349.8	350	350	OK	1546	0.049
Measure K	C	-180	-179.4	-180.6	-180	-180	OK	1546	0.17
Measure K	C	0	0.6	-0.6	0	0	OK	1546	0.17
Measure K	C	400	400.6	399.4	399.9	399.9	OK	1546	0.17
Measure K	C	800	800.8	799.2	799.8	799.8	OK	1546	0.17
Measure K	C	1000	1001	999	999.9	999.9	OK	1546	0.17
Measure K	C	1300	1301.3	1298.7	1299.8	1299.8	OK	1546	0.17
Simulate K	C	-180	-179.4	-180.6	-179.9	-179.9	OK	1546	0.16
Simulate K	C	0	0.6	-0.6	0	0	OK	1546	0.16
Simulate K	C	400	400.6	399.4	400	400	OK	1546	0.16
Simulate K	C	800	800.8	799.2	800	800	OK	1546	0.16
Simulate K	C	1000	1001	999	1000	1000	OK	1546	0.16
Simulate K	C	1300	1301.3	1298.7	1300	1300	OK	1546	0.16
Measure J	C	-200	-199.4	-200.6	-200.3	-200.3	OK	1546	0.15
Measure J	C	0	0.6	-0.6	-0.1	-0.1	OK	1546	0.15
Measure J	C	300	300.6	299.4	299.8	299.8	OK	1546	0.15
Measure J	C	600	600.6	599.4	599.8	599.8	OK	1546	0.15
Measure J	C	900	900.9	899.1	899.8	899.8	OK	1546	0.15
Measure J	C	1200	1201.2	1198.8	1199.8	1199.8	OK	1546	0.15
Simulate J	C	-200	-199.4	-200.6	-199.7	-199.7	OK	1546	0.13
Simulate J	C	0	0.6	-0.6	0.1	0.1	OK	1546	0.13
Simulate J	C	300	300.6	299.4	300.1	300.1	OK	1546	0.13
Simulate J	C	600	600.6	599.4	600.1	600.1	OK	1546	0.13
Simulate J	C	900	900.9	899.1	900.1	900.1	OK	1546	0.13
Simulate J	C	1200	1201.2	1198.8	1200.1	1200.1	OK	1546	0.13

Input range	Eng. Units	Cal Pt	Upper	lower	Initial	Final	Sensitivity	STD	Unc.
							Check		
Measure S	C	1000	1001	999	999	999	OK	1546	0.67
Measure S	C	1150	1151.5	1148.5	1149	1149	OK	1546	0.58
Measure S	C	1300	1301.3	1298.7	1299	1299	OK	1546	0.58
Measure S	C	1450	1451.5	1448.5	1449	1449	OK	1546	0.58
Measure S	C	1600	1601.6	1598.4	1599	1599	OK	1546	0.58
Measure S	C	1750	1751.75	1748.25	1749	1749	OK	1546	0.58
Simulate S	C	1000	1001	999	999.9	999.9	OK	1546	0.67
Simulate S	C	1150	1151.5	1148.5	1150	1150	OK	1546	0.58
Simulate S	C	1300	1301.3	1298.7	1299.9	1299.9	OK	1546	0.58
Simulate S	C	1450	1451.5	1448.5	1450	1450	OK	1546	0.58
Simulate S	C	1600	1601.6	1598.4	1600	1600	OK	1546	0.58
Simulate S	C	1750	1751.75	1748.25	1749.9	1749.9	OK	1546	0.58
Measure N	C	0	0.6	-0.6	-0.3	-0.3	OK	1546	0.2
Measure N	C	200	200.6	199.4	199.7	199.7	OK	1546	0.2
Measure N	C	500	500.6	499.4	499.7	499.7	OK	1546	0.2
Measure N	C	800	800.8	799.2	799.7	799.7	OK	1546	0.2
Measure N	C	1100	1101.1	1098.9	1099.7	1099.7	OK	1546	0.2
Measure N	C	1300	1301.3	1298.7	1299.7	1299.7	OK	1546	0.2
Source N	C	0	0.6	-0.6	0.4	0.4	OK	1546	0.18
Source N	C	200	200.6	199.4	200.2	200.2	OK	1546	0.18
Source N	C	500	500.6	499.4	500.2	500.2	OK	1546	0.18
Source N	C	800	800.8	799.2	800.2	800.2	OK	1546	0.18
Source N	C	1100	1101.1	1098.9	1100.2	1100.2	OK	1546	0.18
Source N	C	1300	1301.3	1298.7	1300.2	1300.2	OK	1546	0.18
Measure T	C	-200	-199.4	-200.6	-199.7	-199.7	OK	1546	0.74
Measure T	C	0	0.6	-0.6	0.2	0.2	OK	1546	0.35
Measure T	C	100	100.6	99.4	100.1	100.1	OK	1546	0.35
Measure T	C	200	200.6	199.4	200.1	200.1	OK	1546	0.35
Measure T	C	300	300.6	299.4	300.1	300.1	OK	1546	0.35
Measure T	C	400	400.6	399.6	400	400	OK	1546	0.35
Source T	C	-200	-199.4	-200.6	-199.9	-199.9	OK	1546	0.73
Source T	C	0	0.6	-0.6	0.1	0.1	OK	1546	0.33
Source T	C	100	100.6	99.4	100.1	100.1	OK	1546	0.33
Source T	C	200	200.6	199.4	200.1	200.1	OK	1546	0.33
Source T	C	300	300.6	299.4	300.1	300.1	OK	1546	0.33
Source T	C	400	400.6	399.6	400	400	OK	1546	0.33
Source mADC lower display	mADC	4	4.0028	3.9972	4.0012	4.0012	OK	1205	0.01
		12	12.0044	11.9956	12.003	12.003	OK	1205	0.01
		24	24.0068	23.9932	24.005	24.005	OK	1205	0.012
Source mVDC lower display	mVDC	0	0.02	-0.02	0.002	0.002	OK	1205	0.0045
		45	45.03	44.97	44.998	44.998	OK	1205	0.01
		100	100.04	99.96	99.996	99.996	OK	1205	0.01
Source VDC lower display	VDC	0	0.002	-0.002	0	0	OK	1205	0.0000045
		5	5.003	4.997	5.0001	5.0001	OK	1205	0.00053
		10	10.004	9.996	10.0003	10.0003	OK	1205	0.0058
Lower display ohm source	Ohms	15	15.1	14.9	15.006	15.006	OK	1205	0.0018
		360	360.1	359.9	359.97	359.97	OK	1205	0.016
		500	500.5	499.5	499.97	499.97	OK	1205	0.016
		1500	1500.5	1499.5	1500.4	1500.4	OK	1205	0.016
		3200	3201	3199	3200.1	3200.1	OK	1205	0.016