



# ILLIANA INSTRUMENTATION

1831 Govert Drive Schererville, IN 46375 Phone (219)227-8788 Fax (219)515-6161

## CERTIFICATE OF CALIBRATION

CUSTOMER:	MISCELLANEOUS DETAILS:
Illiana Instrumentation 1831 Govert Drive Schererville, IN 46375	Date Received 10/9/17 Certification Date: 10/9/17 Recalibration Date: 1/9/18 Cal. Number: 1531-100917 P.O. Number: Location of Calibration: Lab Detailed Results Attached: YES Procedure Used: WI-527

EQUIPMENT CALIBRATED	
MANUFACTURER:	Honeywell Inc.
MODEL:	TVMUSX-88800-240-22-2-030-0U000-000
SERIAL NUMBER:	1126Y192349500001
ITEM NUMBER:	1531
DESCRIPTION:	Video Recorder
CONDITION AS FOUND:	Good, unless otherwise noted on reports

STANDARDS USED/UNCERTAINTIES	
Item 1546 Fluke 525B	
BEST MEASUREMENT UNCERTAINTY:	The expanded (k=2) measurement uncertainty for this test is 1.2 Deg. F.

TEST CONDITIONS	
TEMPERATURE	72 Deg F.
HUMIDITY	57% rH

CERTIFIED BY: Paul Droll TITLE: ISA CCST DATE: 10/9/17  
 APPROVED BY: Laura Grabe TITLE: Deputy Quality Manager DATE: 10/9/17

This certifies that the above equipment was calibrated using appropriate Illiana Instrumentation technical procedures. At planned intervals, Illiana Instrumentation standards are calibrated by comparison to or measurement against standards which are traceable to the SI units through the NIST or other recognized national measurement institutes or international standard bodies. The results in this report relate only to the item(s) calibrated. If so indicated above, detailed calibration results are attached to this certificate. These results are part of this certificate and this certificate shall not be reproduced except in full, without the written approval of Illiana Instrumentation. Any number of factors not under the control of the calibration laboratory may cause the calibration of the above item(s) to drift before the recommended recalibration date. Supporting documentation relative to traceability and technical procedures used is on file and is available for examination upon request and approval of our quality assurance manager. The above uncertainties represent an expanded uncertainty expressed at approximately 95% confidence level using a coverage factor of k=2.

Item No.	1531												
Accuracy	1 Deg F.												
Range	0 to 2200 Deg F. Type K												
Technician	Paul Grolla												
Intentional Offset as found	0												
Intentional Offset as left	0												
Limitations:	None												
Calibration Point	0.00	0	280.0	280.0	760.0	760.0	1240.0	1240.0	1720.0	1720.0	2200.0	2200.0	Sens. Check
	Initial	Final	Initial	Final	Initial	Final	Initial	Final	Initial	Final	Initial	Final	
CH 1	0.40	0	280.2	280.0	760.0	760.0	1240.1	1240.0	1719.9	1720.0	2199.9	2200.0	OK
CH 2	0.00	0	279.7	280.0	759.6	760.0	1239.6	1240.0	1719.5	1720.0	2199.0	2200.0	OK
CH 3	-0.10	0	280.0	280.0	759.9	760.0	1239.9	1240.0	1720.0	1720.0	2199.6	2200.0	OK
CH 4	-0.10	0	279.8	280.0	759.8	760.0	1239.6	1240.0	1719.5	1720.0	2199.5	2200.0	OK
CH 5	0.30	0	280.0	280.0	759.8	760.0	1240.0	1240.0	1719.8	1720.0	2199.4	2200.0	OK
CH 6	0.30	0	280.2	280.0	760.0	760.0	1240.0	1240.0	1719.9	1720.0	2199.6	2200.0	OK
CH 7	0.80	0	280.7	280.0	760.3	760.0	1240.4	1240.0	1720.3	1720.0	2200.0	2200.0	OK
CH 8	0.90	0	280.8	280.0	760.7	760.0	1240.6	1240.0	1720.6	1720.0	2200.3	2200.0	OK
CH 9	0.10	0	280.1	280.0	760.0	760.0	1240.4	1240.0	1720.4	1720.0	2200.6	2200.0	OK
CH 10	0.60	0	280.5	280.0	760.5	760.0	1240.4	1240.0	1720.4	1720.0	2200.2	2200.0	OK
CH 11	-0.20	0	279.9	280.0	759.9	760.0	1240.0	1240.0	1720.1	1720.0	2199.9	2200.0	OK
CH 12	-0.10	0	280.0	280.0	760.0	760.0	1239.9	1240.0	1719.8	1720.0	2199.6	2200.0	OK
CH 13	-0.10	0	280.0	280.0	760.0	760.0	1239.9	1240.0	1719.9	1720.0	2199.7	2200.0	OK
CH 14	0.10	0	280.2	280.0	760.0	760.0	1239.9	1240.0	1719.8	1720.0	2199.6	2200.0	OK
CH 15	0.20	0	280.3	280.0	760.1	760.0	1240.2	1240.0	1720.1	1720.0	2200.3	2200.0	OK
CH 16	0.80	0	280.8	280.0	760.6	760.0	1240.6	1240.0	1720.5	1720.0	2200.4	2200.0	OK