

# CERTIFICATE OF ANALYSIS

CANNON® CERTIFIED VISCOSITY REFERENCE STANDARD					
Viscosity Standard: <b>K400(KU)</b>			Lot Number: <b>15101b</b>		
Certification/Issue Date: <b>03/01/2017</b>			Expiry Date: <b>03/31/2019</b>		
Temperature		Kinematic Viscosity	Dynamic Viscosity	Density	Krebs Unit
° C	° F	mm <sup>2</sup> /s (cSt)	mPa•s (cP)	g/cm <sup>3</sup> (g/mL)	KU
23.00	73.40	1220	1078	0.8832	88.0 K.U.
24.00	75.20	1129	996.1	0.8826	85.8 K.U.
24.50	76.10	1086	958.1	0.8823	84.8 K.U.
25.00	77.00	1045	921.7	0.8820	83.8 K.U.
25.50	77.90	1006	886.9	0.8817	82.7 K.U.
26.00	78.80	968.6	853.7	0.8814	81.8 K.U.

*Tested and certified in the U.S.A.*

This Certificate of Analysis shall not be reproduced, except in full, without the written approval of CANNON Instrument Company.

### USAGE INFORMATION<sup>1</sup>


**Intended Use and Instructions:** This CANNON® Certified Viscosity Reference Standard is intended for but not restricted to the calibration and performance verification of various types of viscometers or density measurement equipment. Consult user's manual and test methods specific to your equipment for operating instructions and procedures.

**Storage and Handling:** This CANNON® Certified Viscosity Reference Standard should be stored in the original container with the lid tightly closed, away from direct light, and at ambient temperatures and normal laboratory conditions. The standard was prepared in accordance with CANNON® Standard Laboratory Operating Procedures to ensure homogeneity and therefore mixing is unnecessary before use and no minimum sample volume is required.

**Composition and Product Safety:** This CANNON® Certified Viscosity Reference Standard is composed of: *Mineral Oil (100%) [CAS#(s) 64742-54-7]*. Consult MSDS for complete product safety information.

**Expiration of Certification:** The certification of this CANNON® Certified Viscosity Reference Standard is valid, within the stated measurement uncertainty, until the expiry date that appears on this certificate, provided the material is stored and handled as stated. This certification is deemed null and void if the standard is modified or contaminated. The shelf life was determined empirically through a historical evaluation of material stability. If substantive technical changes occur to the product which affects the certification before the expiry date, CANNON Instrument Company will contact the purchaser.

 <b>ISO/IEC 17025</b> <b>ACCREDITED</b> <small>Calibration Laboratory CERT#1262.01</small>	 <b>ISO Guide 34</b> <b>ACCREDITED</b> <small>Reference Material Producer CERT#1262.02</small>
<i>The inclusion of the A2LA and ILAC MRA logos does not imply certification/approval of the products calibrated or tested.</i>	
<b>ISO 9001</b> Registered by UL-DQS #10002540 QM	

BOTTLED BY  
**4B** Certification Under  
 Supervision of: 

D.B. Trowbridge, Ph.D.  
 J.T. Mastropiero  
 M.T. Zubler