

1658



CERTIFICATE DATE
January 31, 2018

CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

FLUID	VISCOSITY	TEMPERATURE	LOT NUMBER
5000	4935	25.0	013118

MODEL	(CENTIPOISE)	(°C)	SERIAL NUMBER
LVDV-II+	TEST INSTRUMENT		8545010
MODEL		TEMPERATURE SENSOR	SERIAL NUMBER
1529			A56933

CALIBRATION OF THE ABOVE MENTIONED BROOKFIELD VISCOSITY STANDARD IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE. THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH A BROOKFIELD VISCOMETER THAT HAS BEEN CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

AMETEK Brookfield

11 Commerce Blvd., Middleboro, MA 02346-1031 U.S.A.
TEL 800-628-8139 or 508-946-6200 FAX 508-946-6262
www.brookfieldengineering.com

DAVID J. MOONAY, Ph.D - RHEOLOGY LABORATORY