



CERTIFICATE DATE
February 09, 2018

CERTIFICATE OF VISCOSITY STANDARD CALIBRATION

FLUID	VISCOSITY	TEMPERATURE	LOT NUMBER
1000	1005	25.0	020918

	(CENTIPOISE)	(°C)	
MODEL	TEST INSTRUMENT		SERIAL NUMBER
LVDV-II+			8534910
MODEL	TEMPERATURE SENSOR		SERIAL NUMBER
1502A			B59927

CALIBRATION OF THE ABOVE MENTIONED BROOKFIELD VISCOSITY STANDARD IS ACCURATE TO WITHIN 1% OF ITS STATED VALUE. THE VISCOSITY STANDARD OF THIS LOT NUMBER WAS MEASURED WITH A BROOKFIELD VISCOMETER THAT HAS BEEN CALIBRATED AGAINST A PRIMARY STANDARD, TRACEABLE TO THE NATIONAL INSTITUTE OF STANDARDS & TECHNOLOGY.

AMETEK Brookfield
 11 Commerce Blvd., Middleboro, MA 02346-1031 U.S.A.
 TEL 800-628-8139 or 508-946-6200 FAX 508-946-6262
www.brookfieldengineering.com

D. J. Mooney
 DAVID J. MOONAY, Ph.D - RHEOLOGY LABORATORY SUPERVISOR